

Declaration of Conformity (UKCA)

The Regulations covered by this declaration:

S.I. 2016 No. 1091 Electromagnetic Compatibility Regulations 2016
S.I. 2012 No. 3032 Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment (RoHS) Regulations 2012

Declares that the product(s): **Dakota CX2 & CX4 Ultrasonic Material Thickness Gauge**

Part Number(s): **CX2, CX4**

Product Option(s): **TXC10M0BP-1, TXC10M0CP-4, TXC1M00EP-2, TXC2M25CP-2, TXC2M25EP-2, TXC3M50EP-1, TXC5M00BP-4, TXC5M00CP-4, TXC5M00CP-6, TXC5M00CP-8, TXC5M00EP-3, TXC5M00EP-4, TXC7M50BP-3, TXC7M50CP-4, TXC7M50CP-5, T92024911, T99921325**


This declaration of conformity is issued under the sole responsibility of Elcometer Limited. The products identified above comply with the requirements of the above UK Legislation by meeting the following standards:

EN 61326-1:2013 IEC 61326-1:2012 Class B¹, Group 1² ISM	Electrical equipment for measurement, control and laboratory use – EMC requirements Part 1 General requirements.
EN 61326-2-1:2013¹ IEC 61326-2-1:2012	Electrical equipment for measurement, control and laboratory use — EMC requirements Part 2-1: Particular requirements — Test configurations, operational conditions and performance criteria for sensitive test and measurement equipment for EMC unprotected applications
EN 63000:2018	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

¹ Class B product: Suitable for use in domestic establishments and in establishments directly connected to a low voltage power supply network which supplies buildings used for domestic purposes.

² Group 1 ISM product: Product in which there is intentionally generated and/or used conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.



Signed: 
M. C. Sellars

Manchester, UK
Date: 10th January 2024
Authority: Managing Director